

GT-II 凝膠固化時間測定儀

GT-II Gel Time Tester

GT - II 膠固化時間測定儀按IPC-TM-650試驗方法手冊中的試驗方法2.3.18及國家GB/T16995-1997試驗方法設計製造, 專業用於測試半固化片及樹脂的凝膠化時間。配有必需的各种測試工具。主要元器件均由國外進口, 該系統已廣泛應用在國內外印制電路板與覆銅板製造工業、粉末塗料行業。

The GT-II Gel Time is designed to meet the requirements of IPC-TM-650 Method 2.3.18 and National standards GB/T16995-1997 for determine the gel time of prepreg and mixed resin. The tester is equipped with all necessary tools. The main primary components are import from overseas, this system has been widely put into use in domestic and foreign PCB, CCL factory industry, and the powder coating profession.



技術參數

型號：GT-II 凝膠固化時間測定儀

| | |
|------------|-----------------------------|
| 熱盤控溫精度: | ±0.5°C |
| 熱盤溫度最小讀數: | 0.1°C |
| 工作溫度: | 250°C或300°C |
| 時鐘計時: | 0-999.9秒或更大 |
| 熱盤膠池尺寸: | Φ20.0mm |
| 供電: | 220V, 50Hz, 300W, 單相 |
| 機體尺寸 (mm): | (W x D x H) 310 x 310 x 410 |

Model : GT-II Gel Time Tester

| | |
|-------------------------------------|-----------------------------|
| Hot Platen Temperature Accuracy : | ±0.5°C |
| Hot Platen Temperature Graduation : | 0.1°C |
| Operation Temperature : | 250°C or 300°C |
| Timer : | 0-999.9s or more |
| Hot Platen Resin Pool Size : | Φ20.0mm |
| Power : | 220V, 50Hz, 300W |
| Dimension (mm) : | (W x D x H) 310 x 310 x 410 |

Website : www.eastco-hk.com

EASTCO LIMITED

儀 高 儀 器 有 限 公 司



香港總公司
香港新界荃灣沙咀道26-38號
匯力工業中心2樓26室
Tel : (852)2456 1823
Fax: (852) 2469 1767
E-mail: info@eastco-hk.com

中國公司
儀高南儀器(深圳)有限公司
中國深圳市八卦四路索泰克大廈
3樓V區 郵編: 518029
Tel : (86-755) 8241 3890
Fax: (86-755) 8241 3857
E-mail: shenzhen@eastco-hk.com

上海辦事處
中國上海市滬太路701號
18樓1810室
郵編: 200072
Tel : (86-21) 5653 6311
Fax: (86-21) 5653 6309
E-mail: shanghai@eastco-hk.com